

表 3-2 Minocycline hydrochloride 在血清定量分析之同日確效性試驗

(μ g/ml)	5.000	2.500	1.000	0.500	0.250	0.100
	5.006	2.482	1.020	0.496	0.239	0.108
	5.029	2.437	1.014	0.488	0.270	0.112
	4.977	2.554	0.985	0.491	0.255	0.088
	5.039	2.409	1.025	0.508	0.249	0.119
	4.979	2.545	1.009	0.476	0.237	0.104
	4.990	2.526	1.001	0.474	0.241	0.119
Mean	5.0033	2.4922	1.0090	0.4888	0.2485	0.1083
S.D.	0.0261	0.0597	0.0144	0.0127	0.0125	0.0116
C.V.	0.52%	2.40%	1.43%	2.60%	5.04%	10.71%
error	0.07%	-0.31%	0.90%	-2.23%	-0.60%	8.33%

表 3-3 Minocycline hydrochloride 在血清定量分析之異日確效性試驗

(μ g/ml)	5.000	2.500	1.000	0.500	0.250	0.100
	5.006	2.470	1.058	0.485	0.224	0.108
	4.962	2.583	0.993	0.489	0.228	0.095
	5.015	2.460	1.027	0.499	0.244	0.105
	4.953	2.602	0.987	0.502	0.225	0.080
	5.016	2.464	1.008	0.499	0.255	0.108
	4.985	2.538	0.997	0.473	0.239	0.118
Mean	4.9895	2.5195	1.0117	0.4912	0.2358	0.1023
S.D.	0.0273	0.0636	0.0267	0.0111	0.0124	0.0132
C.V.	0.55%	2.53%	2.64%	2.25%	5.24%	12.89%
error	-0.21%	0.78%	1.17%	-1.77%	-5.67%	2.33%