

表3-4 以一條檢量線回歸 6點 0.1  $\mu\text{g/ml}$  (L.O.Q.) 之確效試驗

測定次數	Minocycline hydrochloride Area	Oxytetracycline Area	Perk area ratio	回歸濃度 ( $\mu\text{g/ml}$ )
1	1488	23038	0.0646	0.0893
2	1451	23716	0.0612	0.0842
3	1639	26147	0.0627	0.0865
4	1512	24342	0.0621	0.0856
5	1433	22210	0.0645	0.0892
6	1282	23117	0.0555	0.0755
mean			0.0618	0.0850
S.D.			0.0034	0.0051
C.V.			5.45%	6.01%

