

表3-7 分析檢品時，安插高、中、低三種標準濃度於檢品中之確效試驗

	LOW 0.1 ($\mu\text{g}/\text{m}$)	MID 1 ($\mu\text{g}/\text{m}$)	HIGH 2.5 ($\mu\text{g}/\text{m}$)
A	0.118	0.997	2.461
B	0.112	1.041	2.449
C	0.112	1.009	2.537
D	0.087	1.016	2.456
E	0.091	1.019	2.526
F	0.081	0.997	2.617
G	0.102	1.127	2.738
H	0.112	0.976	2.352
I	0.101	1.002	2.472
J	0.106	0.970	2.384
K	0.109	0.986	2.551
L	0.115	0.990	2.436
M	0.116	1.012	2.435
N	0.136	0.999	2.497
Mean	0.1070	1.0101	2.4936
S.D.	0.0141	0.0383	0.0980
C.V.	13.17%	3.79%	3.93%
error	7.00%	1.01%	-0.25%

