

Pneumomediastinum, pneumoperitoneum, pneumoretroperitoneum, and subcutaneous emphysema after ultrathin endoscopy



Fig. 1 Endoscopic view of the esophageal cancer.

An 89-year-old woman was admitted for dysphagia over a 2-month period. Transnasal endoscopy with an ultrathin endoscope (Fujinon EG-530N; Fujinon, Wayne, New Jersey, USA) was used. An ulcerative tumor was found in the middle esophagus (● Fig. 1), and the biopsy revealed squamous cell carcinoma. A feeding tube was placed over a guidewire.

A chest computed tomography (CT) scan was performed the next day to evaluate the esophageal cancer. Pneumomediastinum, pneumoperitoneum, pneumoretroperitoneum, and subcutaneous emphysema (● Figs. 2–4) were found, and no fluid collection was seen.

The patient received intravenous antibiotic therapy and nutritional support. An esophageal Niti-S stent (Taewoong Medical, Seoul, Korea) was placed into the patient's esophagus (● Fig. 5).

The patient did not develop clinical signs of sepsis and was placed on a restricted diet 1 week later. She was discharged with the ability to consume semi-solid food and was referred to an oncology clinic.

Perforation of the gastrointestinal tract during endoscopic examination is a rare but serious complication. Perforation with the presence of pneumomediastinum, pneumoperitoneum, pneumoretroperitoneum, and subcutaneous emphysema has been reported in association with endoscopic retrograde cholangiopancreatography [1, 2] and colonoscopy [3]. To our knowledge, this is the first case of such




Fig. 2 Chest computed tomography (CT) revealed subcutaneous emphysema and pneumomediastinum (arrows).




Fig. 3 Abdominal computed tomography (CT) revealed pneumoperitoneum.




Fig. 4 Abdominal computed tomography (CT) revealed pneumoretroperitoneum (arrow).




Fig. 5 Chest radiograph showing the esophageal stent.

complications associated with the ultrathin endoscope [4].

An ultrathin endoscope was used in this case because the patient was elderly and suspected of having esophageal stricture. Indeed, the esophageal lumen was severely stenotic, and the endoscope was passed into the stenotic area after several attempts.

We postulate that the mechanism of perforation in this case was an unrecognized small perforation while manipulating the endoscope through the stenotic area of

the esophageal cancer. Although the ultrathin endoscope is considered safer than a standard endoscope, the clinician should be aware of gastrointestinal perforation, especially in the case of gastrointestinal stricture.

Endoscopic therapy with a metallic stent, as in this case, can be effective to manage complications.

Endoscopy_UCTN_Code_CPL_1AH_2AF

Competing interests: None

顏旭子

H.-H. Yen^{1,2}, Y.-Y. Chen³

¹ Department of Gastroenterology, Changhua Christian Medical Center, Changhua, Taiwan

² Department of Holistic Wellness, MingDao University, Changhua, Taiwan

³ Department of Gastroenterology, China Medical University Hospital, Taichung, Taiwan

References

- 1 Alexiou K, Sakellaris T, Sikalias N et al. Subcutaneous emphysema, pneumomediastinum and pneumoperitoneum after unsuccessful ERCP: a case report. *Cases J* 2009; 2: 120
- 2 Ferrara F, Luigiano C, Billi P et al. Pneumothorax, pneumomediastinum, pneumoperitoneum, pneumoretroperitoneum, and subcutaneous emphysema after ERCP. *Gastrointest Endosc* 2009; 69: 1398 – 1401
- 3 Marwan K, Farmer KC, Varley C, Chapple KS. Pneumothorax, pneumomediastinum, pneumoperitoneum, pneumoretroperitoneum and subcutaneous emphysema following diagnostic colonoscopy. *Ann R Coll Surg Engl* 2007; 89: 548
- 4 ASGE Technology Committee: Rodriguez SA, Banerjee S, Desilets D et al. Ultrathin endoscopes. *Gastrointest Endosc* 2010; 71: 893 – 898

Bibliography

DOI: 10.1055/s-0030-1256057

Endoscopy 2011; 43: E49 – E50

© Georg Thieme Verlag KG Stuttgart · New York · ISSN 0013-726X

Corresponding author

H.-H. Yen, MD

Department of Gastroenterology
Changhua Christian Medical Center

135 Nanshio Street

Changhua 500

Taiwan

Fax: +886-7-7228289

91646@cch.org.tw